

TIQ-F-QC-001 Rev.00

Certificate Of Analysis

Product Name : Buffer Solution pH 8.0 (Phosphate)

Catalog Number : L-1293

Batch Number : 1293-00000000

CAS Number : Not applicable to mixtures

Molecular Weight: Not applicable to mixture

Mol. Formula : Not applicable to mixtures

Expire Date : 24 months

SAMPLE

No	Test	Specifications	Results
1	Appearance	Clear colourless liquid	Clear colourless liquid
2	pH at 25 °C	8.0 ± 0.05	8.01
3	Wt per ml at 20 °C	~ 1.01 g/cm ³	1.005 g/cm ³

LABOTIQ® Buffer Solution pH 8.0 (Phosphate)**Description:**

Clear, colorless phosphate buffer solution intended for calibration and verification of pH measuring instruments.

Method of Analysis :

The pH value was determined at 25°C using a calibrated pH meter.

Traceability Statement:

The reported pH value is traceable to recognized national standards through calibration of the pH measurement system using certified pH buffer reference materials.



SUDIRO S.Si.
Manager QC