

TIQ-F-QC-001 Rev.00

Certificate Of Analysis

Product Name : Buffer Solution pH 9.18

Catalog Number : L-1270

Batch Number : 1270-01260720

CAS Number : Not applicable to mixtures

Molecular Weight: Not applicable to mixture

Mol. Formula : Not applicable to mixtures

Expire Date : Jun, 2028

No	Test	Specifications	Results
1	Appearance	Clear colourless liquid	Clear colourless liquid
2	pH at 25 °C	9.18± 0.02	9.18
3	Wt per ml at 20 °C	~ 1.01 g/cm ³	1.01 g/cm ³

LABOTIQ® Buffer Solution pH 9.18

Material Status :

Calibration Buffer Solution

Method of Value Assignment

The pH value was assigned by potentiometric measurement using a calibrated pH meter equipped with a combined glass electrode at **25.0 ± 0.5°C** after calibration with certified reference buffer solutions.

Traceability Information

Measurement traceable through calibrated instrumentation and certified pH buffer reference materials traceable to NIST or other recognized National Metrology Institutes.



SUDIRO S.Si.
Manager QC